


PDS No. 7881xx	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 02	384 Well Plate, PS, Small Volume™, LoBase			 <b>greiner bio-one</b>
	Greiner Item-No. 7881xx			
Valid for Item-No.:	<b>788101</b>	<b>788161 (sterile)</b>		

1.	Description / Specification	
1.1	Description	PS Microplate, 384 well, Small Volume™, LoBase, solid bottom 788161: sterile
1.2	Dimensions	Plate: length: 127,76 mm (+/- 0,15 mm) width: 85,48 mm (+/- 0,15 mm) curvature: ≤ 150µm
1.3	Volume per well	Total volume: 28 µl (mathematically calculated) Working volume: 4 - 25 µl
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	788101: no 788161: SAL 10 <sup>-3</sup>
1.7	Quality Control	- <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="http://www.gbo.com/bioscience">www.gbo.com/bioscience</a> →Products →Literature →Technical Information →Chemical Resistance of Resins
2.6	Shelf life	788101: - 788161: 5 years after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	10
3.2	Pieces / Box	80
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 01	Date 15 December 2009	Date 16 December 2009	Date 16 December 2009	
Date 21.04.2005	Name S. Kaelberer	Name Dr. R. Heller	Name A. Schulz	