


PDS No. 7681xx	PRODUCT DATA SHEET					Page 1 of 1
Revision 04	Pipettes, CELLSTAR [®] , 50 ml					 greiner bio-one
	Greiner Item-No. 7681xx					
Valid for Item-No.:	768160 (sterile)	768180 (sterile)				

1.	Description / Specification	
1.1	Description	PS Pipette, graduation (scale 1/2 ml), cotton plug, 50 ml, sterile
1.2	Dimensions	Length: 343,0 mm (± 2,0 mm) Ø: 18,5 mm Ø nozzle: 8,0 mm
1.3	Volume	50 ml; volumetric accuracy +/- 2 %
1.4	Material / Resin	Pipette and nozzle: PS (Polystyrene), free of heavy metal Cotton Plug: 100% pure medical cotton, sterile
1.5	Colour	Pipette: clear Cotton Plug: white
1.6	Sterilization	SAL 10 ⁻³
1.7	Quality Control	- Raw Material-Control: physical testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic.
2.2	Temperature range	0°C to +40°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: www.gbo.com/bioscience →Products →Literature →Technical Information →Chemical Resistance of Resins
2.6	Shelf life	5 years after month of production
2.7	Other Information	Fits to all standard pipetting devices

3.	Packaging	768160	768180
3.1	Pieces / Bag	1 "peel off"-packaging (plastic/plastic)	1 "peel off"-packaging (paper/plastic)
3.2	Pieces / Box	100	100
3.3	Lot-No.	YYMMDDXX (year, month, day, index)	YYMMDDXX (year, month, day, index)
3.4	Other Information	Certificate of Quality	Certificate of Quality

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 20 October 2009	Date 21 October 2009	Date 22 October 2009	
Date 16.11.2005	Name S. Kaelberer	Name Dr. L. Marchetti	Name A. Schulz	