


PDS No. 7003xx	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 04	Pasteur Pipette			 <b>greiner bio-one</b>
	Greiner Item-No. 7003xx			
Valid for Item-No.:	<b>700370</b>	<b>700361</b> (sterile)		

1.	Description / Specification	
1.1	Description	Pasteur pipette, capillary tip without graduation 700370: standard 700361: sterile
1.2	Dimensions	Length: 153 mm
1.3	Volume	Working volume: 0,1 ml
1.4	Material / Resin	LDPE (Low Density Polyethylene)
1.5	Colour	Translucent
1.6	Sterilization	700361: SAL 10 <sup>-3</sup> 700370: no
1.7	Quality Control	<u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	-50°C to +80°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="http://www.gbo.com/bioscience">www.gbo.com/bioscience</a> →Products →Literature →Technical Information →Chemical Resistance of Resins
2.6	Shelf life	700361: 5 years after month of production 700370: -
2.7	Other Information	-

3.	Packaging	700361	700370
3.1	Pieces / Bag	25	500
3.2	Pieces / Box	1000	1500
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 03	Date 3 March 2011	Date 4 March 2011	Date 4 March 2011	
Date 30.06.2009	Name S. Kaelberer / S.Jud	Name Dr. L. Marchetti	Name A. Schulz	